



COBI REHAB®
The Positioning Company

XXL-Rehab Clean

XXL-Rehab Clean

Anvendelse
Use

Max. 325 kg

Brugsanvisning
User manual

DK

Bruges til:

- Bariatriske brugere på hospitaler og på plejecentre.

UK

For use by:

- Bariatric users in hospitals and nursing homes.



Brugsanvisning for XXL-Rehab Clean



ADVARSEL

XXL-Rehab Clean må ikke bruges uden grundig instruktion fra uddannet personale eller af personer, der har fået instruktion i brugen af produktet, da forkert tilpasning eller brug kan skade brugeren.

Brugsanvisningen skal læses grundigt før XXL-Rehab Clean tages i brug. Ret specielt opmærksomheden mod Sikkerhedsvejledninger mærket med ⚠.

Cobi Rehab påtager sig intet ansvar for eventuelle skader på produktet eller personskader forårsaget af forkert anvendelse af XXL-Rehab Clean.

Kontakt Cobi Rehab med det samme, hvis der er dele, som er beskadigede eller mangler. Brug aldrig egne erstatningsreservedele.

Cobi Rehab kan ikke holdes ansvarlig for fejl eller ulykker, der opstår efter reparationer, der er lavet uden skriftlig tilladelse fra Cobi Rehab.

Før brug

Boltene til justering af ben og ryg skal monteres korrekt i hullerne på siden af stolen og strammes.

⚠ Bruges toilet- og badestolen, før boltene til justering af ben og ryglæn er blevet monteret og strammet korrekt, kan brugeren falde ud til siden af stolen.

Under brug

⚠ XXL-Rehab Clean er beregnet til badesituationer i vådrumsmiljøer som fx bruserum.

Toilet- og badestolen må ikke bruges i bassiner eller i swimmingpools.

Vær altid opmærksom på usikre løft eller bevægelser, når der forflyttes til eller fra toilet- og badestolen.

Sæderyggen kan indstilles i tre forskellige dybder. Justering af sædedybden påvirker ikke stolens stabilitet.

Anvendelse

XXL-Rehab Clean bruges i forbindelse med toiletbesøg og brusebade af bariatriske brugere, der vejer op til 325 kg.

⚠ Brug ikke toilet- og badestolen til bariatriske brugere over 325 kg. Overskrides maksimal vægten, kan stolen vælte eller gå i stykker.

⚠ Brug kun toilet- og badestolen på hårdt underlag med en maksimal hældning på 3 grader, da stolen ellers kan vælte.



Vedligeholdelse og rengøring

Stolens stel, sæde, armlæn og bækkenspand kan rengøres med vand og et almindeligt rengøringsmiddel.

Sædet kan også tages af, hvis der er behov for desinficering med fx sprit. Sørg for at fastgøre sædet korrekt igen efter rengøring.

Kontrollér regelmæssigt alle bolte og møtrikker. Spænd dem om nødvendigt.

 Manglende kontrol af boltene som anvist kan forårsage funktionsfejl.

Garanti

Cobi Rehab garanterer, at XXL-Rehab Clean er fejlfri, når den leveres. Skulle der mod forventning være mangler eller skade på stolen ved modtagelsen, så kontakt Cobi Rehab med det samme.

Garantiperioden for XXL-Rehab Clean er et år fra købsdato. Fakturaen gælder som garantibevis.

Garantien gælder ikke:

- produkter, hvor serienummer, batchnummer eller lignende er blevet fjernet eller beskadiget væsentligt
- produkter, hvor der er foretaget reparation af uautoriseret personale
- normal vedligeholdelse fx rengøring, tilpasning, smøring og opdatering af udstyr eller dele.

Garantien gælder ikke, hvis Cobi Rehab vurderer, at XXL-Rehab Clean er blevet håndteret på en ureglementeret måde, fx hvis brugervægten overskrides.

Ved skade på grund af uhensigtsmæssigt brug, faktureres der både service og anvendte reservedele.

Garantien, beskrevet ovenover, er den eneste gældende garanti. Cobi Rehab kan ikke holdes ansvarlige for skader af nogen slags.



Sæde med låg



Toiletsæde

Brugervægt max.	325 kg
Total længde	80-101 cm
Total bredde over armlæn	73 / 83 cm
Total højde	79-94 cm
Sædebredde	61 / 71 cm
Sædedybde justerbar	40-52 cm
Sædehøjde justerbar	44-59 cm
Sæderyg højde	37 cm
Bredde toiletåbning	Ø 24 cm, 35 cm lang
Afstand fra gulv til underside af sæde	33-48 cm
Afstand fra gulv til underside af bækkenspand	21-36 cm
Afstand mellem armlæn	67 / 77 cm
Højde sæde til armlæn	26 cm
Vægt	14 / 15 kg
Ramme	Pulverlakeret krom molybdæn stål
Sæde/ryg/armlæn	Polyuretan
Produktets forventede levetid	7 år
ISO - klassifikation	ISO 09 33 07-03



XXL-Rehab Clean er produceret og testet i overensstemmelse med den internationale DS/EN 12182:2012 standard.

Bortskaffelse

Produkter, der ikke længere er i brug, skal bortskaffes på forsvarlig vis. Kontakt kommunen for vejledning om korrekt bortskaffelse.



**PRODUCENT OG
FORHANDLER**

Cobi Rehab
Fuglebækvej 1D
DK-2770 Kastrup
Danmark
Tlf.: +45 70 25 25 22
cobirehab.com

11/2019/CP

Instructions for XXL-Rehab Clean

CAUTION

The bariatric bedside commode/shower chair XXL-Rehab Clean should not be used without proper instruction from trained staff or from persons who have been instructed in the proper use of the product as improper adjustment or use may injure the user.

Please read these instructions thoroughly before using XXL-Rehab Clean. Pay particular attention to the safety instructions marked with .

Cobi Rehab assumes no responsibility for any damage to the product or personal injury caused by improper use of XXL-Rehab Clean.

If components are damaged or missing, please contact Cobi Rehab immediately. Do not use substitute parts.

Cobi Rehab cannot be held liable for defects or accidents caused by repairs carried out without written authorisation from Cobi Rehab.

Before use

The bolts for adjusting the legs and backrest must be properly mounted in the holes on the side of the bariatric bedside commode/shower chair and tightened.

 If the chair is used before the bolts for adjusting the legs and backrest have been properly mounted and tightened, the user may fall off to the side of the chair.

In use

Always be aware of the risk of performing unsafe transfers or movements when moving the user in or out of the bariatric bedside commode/shower chair.

The backrest can be adjusted to three different depths. This does not affect the stability of the chair.

The surface temperature of the bariatric bedside commode/shower chair can increase or decrease when exposed to external sources of heat or cold.

Ensure that all four legs are level to avoid instability problems.

Use

XXL-Rehab Clean is used by bariatric users weighing up to 325 kg during lavatory use and in shower situations.

 If the max. weight limit is exceeded the chair may tip or break.

 Only use bariatric bedside commode/shower chair on a firm floor with a maximum slope of 3 degrees, as the chair may otherwise tip.

 The shower chair is only suitable for showering in wet rooms such as shower rooms. Do not use the shower chair in pools or swimming pools.



Maintenance and cleaning

The chair frame, seat, armrests and the commode bucket may be cleaned with water and a household detergent.

The seat is also removable if the need for disinfection with e.g. white spirit is required. Make sure to fasten the seat properly again after cleaning.

Regularly check all bolts and nuts and retighten if required.

 Non-compliance with the bolt inspection guidelines may lead to malfunction.

Warranty

Cobi Rehab warrants that the XXL-Rehab Clean is without defects on delivery. If, against all expectations, the bariatric bedside commode/shower chair is subject to deficiencies or defects on receipt, please contact Cobi Rehab immediately.

The warranty period for the XXL-Rehab Clean is one year from the date of purchase. The invoice is your certificate of warranty.

The warranty does not apply to:

- products on which the serial number, batch number or the like have been removed or significantly damaged
- which have been repaired by unauthorized personnel
- normal maintenance such as cleaning, adjustment, lubrication and updating of equipment or parts.

This warranty does not cover Clean when Cobi Rehab assesses that XXL-Rehab Clean has been improperly handled, e.g. if the maximum user weight has been exceeded.

For damage caused by improper operation, an invoice will be issued for both the service and the spare parts used.

The above warranty is the only warranty offered. Cobi Rehab cannot be held liable for damage or injury of any kind.



Seat with lid



Split toilet seat

Maximum user weight	325 kg
Total length	80-101 cm
Total outer width on armrests	73 / 83 cm
Total height	79-94 cm
Seat width	61 / 71 cm
Seat depth, adjustable	40-52 cm
Seat height, adjustable	44-59 cm
Height of backrest	37 cm
Width toilet opening	Diameter 24 cm, 35 cm long
Distance from floor to seat underside	33-48 cm
Distance from floor to bucket underside	21-36 cm
Distance between armrests	69 / 79 cm
Seat to armrest	26 cm
Weight	14 / 15 kg
Frame	Powder-lacquered chromium molybdenum steel
Seat / back / armrests	Polyurethane
ISO - classification	ISO 09 33 03
Expected product life service	7 years



XXL-Rehab Clean is manufactured and tested in accordance with the international DS/EN 12182:2012 standard.

Disposal

Products which are no longer in use, must be responsibly disposed of. Contact the relevant local authorities for guidance on correct disposal.



MANUFACTURER AND DEALER

Cobi Rehab
Fuglebækvej 1D
DK-2770 Kastrup
Denmark
Tel.: +45 70 25 25 22
cobirehab.com

11/2019/CP



COBI REHAB[®]

The Positioning Company

COBI REHAB

Fuglebækvej 1D
DK-2770 Kastrup
Denmark
Tel. +45 7025 2522
cobirehab.com